



NOTE: PUMP WILL NOT OPERATE IF THE SAMPLE TRAIN IS COLD OR THE N2/CAL SVS ARE IN A POSITION TO ALLOW GAS TO FLOW. THIS PROTECTS THE OPTICS INSIDE THE MULTI-GAS FROM WATER DAMAGE, AVOIDS WASTING CAL/ZERO GAS AND REDUCES THE CHANCES FOR OPERATOR ERROR.

*WITH SAFETY INTERLOCK AND INCOMING SAMPLE TEMP SENSOR

ATMO-SEAL ENG., INC.
1091 WHEATON AVE, TROY MI 48063
WWW.ATMOSEAL-ENG.COM

PART #	DHS-M46 FLOW SCHEMATIC		
DOC #	DHSM46FL		
ENG BY	SCALE	NTS	
CHK BY	DATE	10/09	
REV #	DATE		